

Certificate



Certificate no.

CU 72081031 01

License Holder:

Lode B.V. Medical Technology
Zernikepark 16
9747 AN Groningen
Netherlands

Manufacturing Plant:

Lode B.V. Medical Technology
Zernikepark 16
9747 AN Groningen
Netherlands

Test report no.: USA-DS 30793714 001

Client Reference: Jaap Noordmans

Tested to: UL 60335-1:2004
ASTM F1250-00
CAN/CSA E335-1/3E

Certified Product: Ergometer

License Fee - Units

Model Designation: Angio Series

7

Rated Voltage: AC 115V, 50/60Hz
Rated Power: 130VA
Protection Class: I

Special Remarks: This equipment was also evaluated to
UL 60601-1:2003 and CAN/CSA C22.2 No. 601-M90.

7

Appendix: 1, 1-2

Licensed Test mark:



Signature

Dipl.-Ing. M. Raap
QA Certification Officer

Date of Issue

(day/mo/yr)

07/07/2008